

Info Sheet INFO 110.E-ed.G

Date of issue: 01/03/17 Replaces version from: 01/03/07

Page 1/1

NEW PROSTHESIS WITH DAMAGED FEMALES (M3)

If a new prosthesis must be fabricated in case of abutment retainers with damaged females, the following procedure must be followed.

CHAIRSIDE PROCEDURE

- 1. Use the **H 30** accessory to determine the diameter of the spring pin for the new prosthesis (see INFO 067).
- 2. Adjust the shape of the impression tray.
- 3. Assemble the **H 14** impression tool with the selected oversized spring pin and the space maintainer. Place this combination in the female in the mouth.
- 4. The accessory combination must be absolutely stable and secure in the female. If not, activate the spring pin (see INFO 059).
- 5. Block out the undercuts of the female with soft wax.
- 6. Take impression with a custom-made impression tray.

LABORATORY PROCEDURE

- Replace the spring pin used during impression-taking with a 694 C normal spring pin.
 Assemble the H 14 impression tool including the 694 C normale spring pin with a 694 B space maintainer and the H 13 model accessory.
- 2. Place this combination in the impression.
- 3. Pour a model in stone which incorporates the model analogues.
- 4. Make a new secondary construction.
- After fabrication of the new prosthesis, replace the 694 C normal spring pins with the oversized spring pins used during impression-taking and secure them with CEKA BOND (see INFO 069).

Side effects, warnings and precautions

- The attachments are intended for single use.
- The products are non-sterile.
- There is a risk of poor fit when patient conditions change.
- Bacterial adhesion can be avoided by applying hygiene measures.
- Inappropriate use or bad manufacturing can lead to premature wear of the attachments.
- The functionality of the attachments will be adversely affected by traumas such as grinding and bruxism.
- For the purpose of traceability we advise you to record the lot number of the applied products in the patient file.
- Do not heat items containing titanium.
- Do not use items containing nickel in case of nickel allergy.
- The accessories RE H 79 and H 35 must be used outside the mouth.